

CE



09

DECLARATION OF CONFORMITY

We,

Dwyer Instruments, Inc.
P.O. Box 373
Michigan City, IN 46361, USA
(219) 897-8868

declare under our sole responsibility that our Series TTE to which this declaration relates are in Conformity with the following standards:

- EN 61000-4-2 (2001) Electromagnetic Compatibility (EMC) – Part 4-2: Testing Measurement Techniques – Electrostatic Discharge Immunity Test
- EN 61000-4-3 (2006) Electromagnetic Compatibility (EMC) – Part 4-3: Testing Measurement Techniques – Radio Frequency, Electromagnetic Field Immunity Test
- EN 61000-4-4 (2004) Electromagnetic Compatibility (EMC) – Part 4-4: Testing Measurement Techniques – Electrical Fast Transient/Burst Immunity Test
- EN 61000-4-5 (2005) Electromagnetic Compatibility (EMC) – Part 4-4: Testing Measurement Techniques – Surge Immunity Test
- EN 61000-4-6 (2006) Electromagnetic Compatibility (EMC) – Part 4-6: Testing Measurement Techniques – Immunity To Conducted Disturbances, Induced By Frequency Fields
- EN55011 (2003) Limits and methods of measurement of radio disturbance characteristics of industrial, scientific and medical (ISM) radio-frequency equipment
- EN61326-1(2003) Electrical equipment for measurement, control and laboratory use. EMC requirements

Per the provisions of the Electromagnetic Compatibility Directive 2004/108/EC

The authorized representative located within the Community is:

Dwyer Instruments Ltd
Unit 16, The Wye Estate, London Road
High Wycombe, Bucks HP11 1LH-U.K.
(+44) (0) 1494 461707

On behalf of Dwyer Instruments, Inc.

A handwritten signature in black ink, appearing to read "Ling Pott", written over a horizontal line.

Senior Project Engineer

Michigan City, Indiana, USA
06/30/09

